


DECLARATION OF DISINFECTANT EFFECTIVENESS (ALSO VIRUCIDAL AGAINST SARS-CoV-2) Steam Disinfection Devices without contact / with contact

The company **Polti Spa** declares and certifies that the products marked **“SDD” - Steam Disinfection Devices**, used in the correct operating mode, **have bactericidal, sporicidal, fungicidal, levurocidal and virucidal disinfectant efficacy - also against SARS-CoV-2.**

The effectiveness has been tested by independent third-party laboratories in unclean conditions with substances that interfere with human / medical health in compliance with the **AFNOR NF T72-110 standard - March 2019**



Bench test for the evaluation of steam devices valid for "contact" and "without contact" tests
 C4 / C5 / C6 = areas contaminated with microorganisms
 The passing speed complies with the standard: 10 cm / s
 The distance for the "without contact" test complies with the standard: 5 mm
 In the "contact" test, a single step was performed for all rows of specimens. In the "without contact" test, one pass was performed for each row of specimens.

PRODUCTS TESTED IN THE CORRECT OPERATING MODES PROVIDED FOR BY THE STANDARD OBTAIN THE FOLLOWING REDUCTIONS:

- ≥5 Log - Bacteria - Bactericidal activity
- ≥4 Log - Virus - Virucidal Activity
- ≥4 Log - Fungi - Fungicidal activity
- ≥4 Log - Yeasts - Levurocidal activity
- ≥3 Log - Bacterial Spores - Sporicidal Activity



**STEAM
DISINFECTION
DEVICE**
(ref. AFNOR NF T72-110)

